

EQUIPMENT INFORMATION

BASIC INFORMATION

- ☒ Data based on maximum power source capability
- ☐ Data based on worst case setting/program
- ☐ Data based on multiple settings / programs

COMPLIANCE SUMMARY

| | | |
|---|---|--|
| Compliance with ELVs for health effects at standardized configurations | <input checked="" type="checkbox"/> yes | <input type="checkbox"/> no |
| Compliance with ELVs for sensory effects at standardized configurations | <input checked="" type="checkbox"/> yes | <input type="checkbox"/> no |
| All values below the Action Levels at the standardized configurations | <input type="checkbox"/> yes | <input checked="" type="checkbox"/> no |

NON-THERMAL EFFECTS

| | HEAD | | TRUNK | LIMB HAND | LIMB THIGH |
|-----------------------------|-----------------|----------------|----------|-----------|------------|
| | SENSORY EFFECTS | HEALTH EFFECTS | | | |
| Standardized distance | 10.00 cm | 10.00 cm | 10.00 cm | 3.00 cm | 3.00 cm |
| EI at standardized distance | 0.21 | 0.21 | 0.34 | 0.19 | 0.44 |
| Required minimum distance | 0.50 cm | 0.50 cm | 1.00 cm | 1.00 cm | 0.50 cm |

| | |
|--|----------|
| Distance where all EIs fall below 20: % | 22.00 cm |
| Distance where compliance with general public limits is reached: | 1.58 m |